

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040619\
 Data File : VU030785.D
 Acq On : 06 Apr 2019 08:41
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: Apr 08 02:20:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 01:03:28 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	457647	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	460937	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	185735	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	197108	48.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.50%
7) Chloroethane-d5	1.68	69	160041	51.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.24%
11) 1,1-Dichloroethene-d2	2.27	63	282161	38.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.38%
21) 2-Butanone-d5	4.17	46	358137	110.61	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	110.61%
24) Chloroform-d	4.64	84	305104	48.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.78%
26) 1,2-Dichloroethane-d4	5.31	65	223764	52.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.10%
32) Benzene-d6	5.34	84	671621	49.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.82%
36) 1,2-Dichloropropane-d6	6.32	67	233714	50.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.36%
41) Toluene-d8	7.56	98	582318	48.82	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.64%
43) trans-1,3-Dichloropropene-	7.85	79	95803	45.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.12%
47) 2-Hexanone-d5	8.31	63	208458	98.90	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.90%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	311186	50.03	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.06%
64) 1,2-Dichlorobenzene-d4	11.86	152	197633	54.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040619\
Data File : VU030785.D
Acq On : 06 Apr 2019 08:41
Operator : JC/SP
Sample : VIBLK40
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Apr 08 02:20:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 08 01:03:28 2019
Response via : Initial Calibration

