

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040620\
 Data File : VU037584.D
 Acq On : 06 Apr 2020 11:55
 Operator : JC/MD
 Sample : VU0406WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK47

Quant Time: Apr 07 01:11:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 07 00:48:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	182687	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	173829	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	80260	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	56657	38.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.38%
7) Chloroethane-d5	1.93	69	46808	39.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.54%
11) 1,1-Dichloroethene-d2	2.59	63	79670	31.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.88%
21) 2-Butanone-d5	4.66	46	97265	87.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.22%
24) Chloroform-d	5.11	84	99994	40.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.00%
26) 1,2-Dichloroethane-d4	5.74	65	67580	40.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.44%
32) Benzene-d6	5.76	84	211402	41.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.94%
36) 1,2-Dichloropropane-d6	6.72	67	72274	42.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.26%
41) Toluene-d8	7.93	98	193829	41.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.58%
43) trans-1,3-Dichloropropene-	8.20	79	36760	39.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.86%
47) 2-Hexanone-d5	8.66	63	75868	84.46	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.46%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	100313	41.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.78%
64) 1,2-Dichlorobenzene-d4	12.21	152	66551	40.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040620\
Data File : VU037584.D
Acq On : 06 Apr 2020 11:55
Operator : JC/MD
Sample : VU0406WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK47

Quant Time: Apr 07 01:11:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 07 00:48:05 2020
Response via : Initial Calibration

