

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040620\
 Data File : VU037590.D
 Acq On : 06 Apr 2020 15:08
 Operator : JC/MD
 Sample : L2158-08 10X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0F05

Quant Time: Apr 07 01:43:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040320WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 07 00:48:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	171401	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	161488	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	77148	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	57010	40.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.90%
7) Chloroethane-d5	1.91	69	46739	41.79	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.58%
11) 1,1-Dichloroethene-d2	2.58	63	81095	34.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	69.32%
21) 2-Butanone-d5	4.66	46	97876	93.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.55%
24) Chloroform-d	5.10	84	105229	44.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.74%
26) 1,2-Dichloroethane-d4	5.74	65	69817	44.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.68%
32) Benzene-d6	5.76	84	214371	45.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.62%
36) 1,2-Dichloropropane-d6	6.72	67	72911	46.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.58%
41) Toluene-d8	7.92	98	196215	45.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.00%
43) trans-1,3-Dichloropropene-	8.20	79	37006	43.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.54%
47) 2-Hexanone-d5	8.65	63	80726	96.74	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.74%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	109473	48.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.24%
64) 1,2-Dichlorobenzene-d4	12.20	152	72425	46.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.02%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.59	96	3154	3.440 ug/L	# 1
20) cis-1,2-Dichloroethene	4.71	96	11773	10.355 ug/L	90
34) Trichloroethene	6.57	95	166280	150.288 ug/L	99
46) Tetrachloroethene	8.57	164	6591	8.872 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040620\
Data File : VU037590.D
Acq On : 06 Apr 2020 15:08
Operator : JC/MD
Sample : L2158-08 10X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0F05

Quant Time: Apr 07 01:43:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040320WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 07 00:48:05 2020
Response via : Initial Calibration

