

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
 Data File : VU042948.D
 Acq On : 06 Apr 2021 12:46
 Operator : SY/MD
 Sample : M1909-02MEDL 100X
 Misc : 6.34g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FS-SB-10-26-27MEDL

Quant Time: Apr 07 01:41:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	127783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	207366	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	198123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	106441	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	102874	50.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.00%	
35) Dibromofluoromethane	5.298	113	71454	46.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.34%	
50) Toluene-d8	7.906	98	259949	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	10.639	95	102845	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
Target Compounds					Qvalue	
40) Benzene	5.784	78	46388	7.56	ug/l	98
52) Toluene	7.976	92	80124	21.36	ug/l	99
67) Ethyl Benzene	9.578	91	221317	30.63	ug/l	99
68) m/p-Xylenes	9.700	106	223892	84.08	ug/l	99
69) o-Xylene	10.108	106	90561	35.14	ug/l	99
73) Isopropylbenzene	10.491	105	18894	2.80	ug/l	99
78) n-propylbenzene	10.912	91	30863	3.89	ug/l	99
80) 1,3,5-Trimethylbenzene	11.095	105	68853	11.78	ug/l	100
84) 1,2,4-Trimethylbenzene	11.475	105	205950	34.89	ug/l	97
95) Naphthalene	14.095	128	1171317	156.52	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
Data File : VU042948.D
Acq On : 06 Apr 2021 12:46
Operator : SY/MD
Sample : M1909-02MEDL 100X
Misc : 6.34g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FS-SB-10-26-27MEDL

Quant Time: Apr 07 01:41:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

