

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
 Data File : VU042955.D
 Acq On : 06 Apr 2021 15:26
 Operator : SY/MD
 Sample : M1894-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 0156-AMENDED-CP-COMP

Quant Time: Apr 07 01:43:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.385	168	124756	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	200377	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	191474	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	93797	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	99767	50.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.32%	
35) Dibromofluoromethane	5.298	113	60449	40.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.84%	
50) Toluene-d8	7.906	98	251066	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
62) 4-Bromofluorobenzene	10.639	95	92133	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
16) Acetone	2.642	43	31170	18.35	ug/l	98
20) Methylene Chloride	3.057	84	3112	2.17	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
Data File : VU042955.D
Acq On : 06 Apr 2021 15:26
Operator : SY/MD
Sample : M1894-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
0156-AMENDED-CP-COMP

Quant Time: Apr 07 01:43:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

