

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
 Data File : VU042958.D
 Acq On : 06 Apr 2021 16:34
 Operator : SY/MD
 Sample : M1951-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB-2021-04-05_SW

Quant Time: Apr 07 01:44:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 25 05:42:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	120590	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.260	114	193382	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	187422	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	91834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.713	65	97767	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
35) Dibromofluoromethane	5.298	113	67762	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
50) Toluene-d8	7.906	98	246934	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	10.639	95	90838	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	

Target Compounds	Qvalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
Data File : VU042958.D
Acq On : 06 Apr 2021 16:34
Operator : SY/MD
Sample : M1951-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB-2021-04-05_SW

Quant Time: Apr 07 01:44:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 25 05:42:00 2021
Response via : Initial Calibration

