



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040621\
 Data File : VU042942.D
 Acq On : 06 Apr 2021 09:24
 Operator : SY/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U032421W.M
 Title : SW846 8260
 Last Update : Thu Mar 25 05:42:00 2021



AutoFind: Scans 2892, 2893, 2894; Background Corrected with Scan 2880

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.7	5871	PASS
75	95	30	60	58.2	13842	PASS
95	95	100	100	100.0	23792	PASS
96	95	5	9	7.1	1701	PASS
173	174	0.00	2	0.6	117	PASS
174	95	50	100	86.9	20675	PASS
175	174	5	9	8.1	1681	PASS
176	174	95	101	97.7	20195	PASS
177	176	5	9	6.7	1361	PASS