

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040623\
 Data File : VU053586.D
 Acq On : 06 Apr 2023 11:37
 Operator : JC/MD
 Sample : VU0406MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK071

Quant Time: Apr 07 03:45:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 01 04:51:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	255798	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	235431	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	117711	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	101072	50.744	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.480%
7) Chloroethane-d5	1.909	69	82616	53.796	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.600%
11) 1,1-Dichloroethene-d2	2.565	63	118517	38.439	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.880%
21) 2-Butanone-d5	4.613	46	105406	116.884	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.880%
24) Chloroform-d	5.057	84	160170	48.624	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.240%
26) 1,2-Dichloroethane-d4	5.697	65	101424	52.106	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.220%
32) Benzene-d6	5.722	84	333946	52.549	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.100%
36) 1,2-Dichloropropane-d6	6.687	67	103763	53.420	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.840%
41) Toluene-d8	7.893	98	312854	49.532	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.060%
43) trans-1,3-Dichloroprop...	8.176	79	49057	48.927	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.860%
47) 2-Hexanone-d5	8.626	63	79201	119.473	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	119.470%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	146447	53.492	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.980%
66) 1,2-Dichlorobenzene-d4	12.188	152	119666	48.703	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.400%
Target Compounds						
70) 1,2,4-trichlorobenzene	13.841	180	3906	1.564	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040623\
Data File : VU053586.D
Acq On : 06 Apr 2023 11:37
Operator : JC/MD
Sample : VU0406MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK071

Quant Time: Apr 07 03:45:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 01 04:51:17 2023
Response via : Initial Calibration

