

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
 Data File : VU030796.D
 Acq On : 06 Apr 2019 13:37
 Operator : JC/SP
 Sample : K2215-20
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 08 05:18:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 05:01:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	451169	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	441028	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	186263	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	205918	51.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.26%
7) Chloroethane-d5	1.68	69	163893	53.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.24%
11) 1,1-Dichloroethene-d2	2.27	63	291711	40.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.10%
21) 2-Butanone-d5	4.17	46	372895	116.82	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.82%
24) Chloroform-d	4.64	84	308023	50.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.14%
26) 1,2-Dichloroethane-d4	5.31	65	232202	55.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.62%
32) Benzene-d6	5.33	84	663049	51.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.00%
36) 1,2-Dichloropropane-d6	6.32	67	236044	53.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.00%
41) Toluene-d8	7.56	98	570016	49.95	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.90%
43) trans-1,3-Dichloropropene-	7.85	79	104751	52.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.12%
47) 2-Hexanone-d5	8.31	63	222977	110.57	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.57%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	318591	53.53	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.06%
64) 1,2-Dichlorobenzene-d4	11.85	152	193074	52.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.42%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
Data File : VU030796.D
Acq On : 06 Apr 2019 13:37
Operator : JC/SP
Sample : K2215-20
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 08 05:18:07 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 08 05:01:06 2019
Response via : Initial Calibration

