

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
 Data File : VU030819.D
 Acq On : 06 Apr 2019 22:49
 Operator : JC/SP
 Sample : K2218-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 08 07:31:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 07:03:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	497598	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	492074	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	193058	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	228319	51.40	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.80%
7) Chloroethane-d5	1.68	69	180600	53.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.14%
11) 1,1-Dichloroethene-d2	2.27	63	321748	40.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.10%
21) 2-Butanone-d5	4.17	46	375948	106.79	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.79%
24) Chloroform-d	4.64	84	333191	49.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.30	65	245941	53.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.24%
32) Benzene-d6	5.34	84	731898	50.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.90%
36) 1,2-Dichloropropane-d6	6.32	67	251362	51.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.12%
41) Toluene-d8	7.56	98	624344	49.03	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.06%
43) trans-1,3-Dichloropropene-	7.85	79	101315	45.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.26%
47) 2-Hexanone-d5	8.31	63	237093	105.37	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.37%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	332291	50.04	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.08%
64) 1,2-Dichlorobenzene-d4	11.85	152	206298	54.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.66%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040719\
Data File : VU030819.D
Acq On : 06 Apr 2019 22:49
Operator : JC/SP
Sample : K2218-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 08 07:31:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 08 07:03:53 2019
Response via : Initial Calibration

