

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
 Data File : VU030794.D
 Acq On : 06 Apr 2019 12:52
 Operator : JC/SP
 Sample : VU0407MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK47

Quant Time: Apr 08 05:13:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 05:01:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	482112	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	476334	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	201326	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	226382	52.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	105.20%
7) Chloroethane-d5	1.68	69	176357	53.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	107.98%
11) 1,1-Dichloroethene-d2	2.27	63	316589	40.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.34%
21) 2-Butanone-d5	4.17	46	382781	112.22	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.22%
24) Chloroform-d	4.64	84	334272	50.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.70%
26) 1,2-Dichloroethane-d4	5.30	65	239890	53.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.96%
32) Benzene-d6	5.33	84	712278	51.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.44%
36) 1,2-Dichloropropane-d6	6.32	67	251219	52.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.44%
41) Toluene-d8	7.56	98	622221	50.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.96%
43) trans-1,3-Dichloropropene-	7.85	79	110888	51.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	102.06%
47) 2-Hexanone-d5	8.31	63	226384	103.94	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.94%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	324532	50.49	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.98%
64) 1,2-Dichlorobenzene-d4	11.86	152	211135	53.33	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.66%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
Data File : VU030794.D
Acq On : 06 Apr 2019 12:52
Operator : JC/SP
Sample : VU0407MBL01
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK47

Quant Time: Apr 08 05:13:02 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 08 05:01:06 2019
Response via : Initial Calibration

