

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
 Data File : VU030818.D
 Acq On : 06 Apr 2019 22:26
 Operator : JC/SP
 Sample : VU0407WBL02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK49

Quant Time: Apr 08 07:18:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Apr 08 07:03:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	481437	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	471819	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	188847	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	221375	51.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.02%
7) Chloroethane-d5	1.68	69	176986	54.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.52%
11) 1,1-Dichloroethene-d2	2.27	63	311001	40.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.02%
21) 2-Butanone-d5	4.17	46	373463	109.64	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.64%
24) Chloroform-d	4.64	84	333051	50.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.46%
26) 1,2-Dichloroethane-d4	5.30	65	241232	53.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.70%
32) Benzene-d6	5.34	84	700952	50.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.78%
36) 1,2-Dichloropropane-d6	6.32	67	245790	52.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.14%
41) Toluene-d8	7.56	98	603270	49.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.82%
43) trans-1,3-Dichloropropene-	7.85	79	102539	47.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.28%
47) 2-Hexanone-d5	8.31	63	217057	100.61	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.61%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	321461	50.49	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.98%
64) 1,2-Dichlorobenzene-d4	11.86	152	199180	53.63	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040719\
Data File : VU030818.D
Acq On : 06 Apr 2019 22:26
Operator : JC/SP
Sample : VU0407WBL02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK49

Quant Time: Apr 08 07:18:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Apr 08 07:03:53 2019
Response via : Initial Calibration

