

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040721\
 Data File : VU042977.D
 Acq On : 07 Apr 2021 12:30
 Operator : SY/MD
 Sample : VU0407MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK307

Quant Time: Apr 08 08:56:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM040721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 08 08:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	192226	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	180029	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	85761	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	55082	50.30	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.60%
7) Chloroethane-d5	1.909	69	46074	49.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.40%
11) 1,1-Dichloroethene-d2	2.572	63	91362	35.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.08%
21) 2-Butanone-d5	4.626	46	150168	97.10	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.10%
24) Chloroform-d	5.070	84	123707	45.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	5.710	65	96840	47.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.92%
32) Benzene-d6	5.735	84	233034	50.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.90%
36) 1,2-Dichloropropane-d6	6.697	67	80523	51.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.42%
41) Toluene-d8	7.906	98	218037	47.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.66%
43) trans-1,3-Dichloroprop...	8.185	79	40032	48.47	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.94%
47) 2-Hexanone-d5	8.639	63	97186	99.17	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.17%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	123319	46.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.62%
66) 1,2-Dichlorobenzene-d4	12.201	152	88513	51.31	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.62%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040721\
Data File : VU042977.D
Acq On : 07 Apr 2021 12:30
Operator : SY/MD
Sample : VU0407MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK307

Quant Time: Apr 08 08:56:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM040721WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 08 08:47:10 2021
Response via : Initial Calibration

