

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040721\
 Data File : VU042978.D
 Acq On : 07 Apr 2021 12:54
 Operator : SY/MD
 Sample : VU0407WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK308

Quant Time: Apr 08 08:56:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM040721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 08 08:47:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	180464	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	170453	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	83518	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	48667	47.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.68%
7) Chloroethane-d5	1.903	69	43560	49.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	99.10%
11) 1,1-Dichloroethene-d2	2.569	63	85050	35.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.48%
21) 2-Butanone-d5	4.626	46	141714	97.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.60%
24) Chloroform-d	5.070	84	115498	45.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.30%
26) 1,2-Dichloroethane-d4	5.710	65	90529	47.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.52%
32) Benzene-d6	5.736	84	216051	49.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.78%
36) 1,2-Dichloropropane-d6	6.700	67	73397	49.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.60%
41) Toluene-d8	7.906	98	206807	47.91	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.82%
43) trans-1,3-Dichloroprop...	8.186	79	37817	48.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.72%
47) 2-Hexanone-d5	8.639	63	92178	99.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.34%
56) 1,1,2,2-Tetrachloroeth...	10.764	84	121694	48.27	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.54%
66) 1,2-Dichlorobenzene-d4	12.202	152	84869	50.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.04%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040721\
Data File : VU042978.D
Acq On : 07 Apr 2021 12:54
Operator : SY/MD
Sample : VU0407WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK308

Quant Time: Apr 08 08:56:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM040721WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 08 08:47:10 2021
Response via : Initial Calibration

