

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU040819\
 Data File : VU030859.D
 Acq On : 07 Apr 2019 16:09
 Operator : JC/SP
 Sample : K2283-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZ92

Quant Time: Apr 08 04:37:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 03:35:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	246621	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	255928	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	101771	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	113254	6.01	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.20%
7) Chloroethane-d5	1.68	69	93103	6.13	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	122.60%
11) 1,1-Dichloroethene-d2	2.27	63	166250	4.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.20%
20) 2-Butanone-d5	4.19	46	310937	64.03	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	128.06%
24) Chloroform-d	4.64	84	162868	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
26) 1,2-Dichloroethane-d4	5.31	65	106752	6.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.80%
32) Benzene-d6	5.33	84	357442	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.00%
36) 1,2-Dichloropropane-d6	6.33	67	124534	5.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.20%
41) Toluene-d8	7.56	98	289503	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	7.85	79	41375	4.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.60%
46) 2-Hexanone-d5	8.31	63	194437	53.27	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.54%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	96689	5.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	119.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	100399	5.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU040819\
Data File : VU030859.D
Acq On : 07 Apr 2019 16:09
Operator : JC/SP
Sample : K2283-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZ92

Quant Time: Apr 08 04:37:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 03:35:14 2019
Response via : Initial Calibration

