

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU040819\
 Data File : VU030868.D
 Acq On : 07 Apr 2019 20:10
 Operator : JC/SP
 Sample : K2284-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZ99

Quant Time: Apr 08 05:05:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 04:40:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	231974	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	238767	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	94832	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	108285	6.11	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.20%
7) Chloroethane-d5	1.68	69	88203	6.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	123.40%
11) 1,1-Dichloroethene-d2	2.27	63	155456	4.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.60%
20) 2-Butanone-d5	4.19	46	286097	62.63	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	125.26%
24) Chloroform-d	4.64	84	152439	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
26) 1,2-Dichloroethane-d4	5.31	65	100596	6.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	126.00%
32) Benzene-d6	5.33	84	341338	5.37	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.40%
36) 1,2-Dichloropropane-d6	6.32	67	112602	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.56	98	271399	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
43) trans-1,3-Dichloropropene-	7.85	79	37417	4.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.60%
46) 2-Hexanone-d5	8.31	63	183456	53.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.74%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	91733	6.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	121.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	92779	5.85	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU040819\
Data File : VU030868.D
Acq On : 07 Apr 2019 20:10
Operator : JC/SP
Sample : K2284-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZ99

Quant Time: Apr 08 05:05:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 04:40:32 2019
Response via : Initial Calibration

