

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU040819\
 Data File : VU030880.D
 Acq On : 08 Apr 2019 01:01
 Operator : JC/SP
 Sample : K2284-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZB1

Quant Time: Apr 08 05:56:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 04:40:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	231762	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	244277	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	94629	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	110885	6.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	125.20%
7) Chloroethane-d5	1.68	69	93621	6.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	131.20%#
11) 1,1-Dichloroethene-d2	2.27	63	160189	4.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.40%
20) 2-Butanone-d5	4.19	46	301532	66.07	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	132.14%#
24) Chloroform-d	4.64	84	157344	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.80%
26) 1,2-Dichloroethane-d4	5.31	65	107021	6.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	134.20%#
32) Benzene-d6	5.33	84	357232	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.00%
36) 1,2-Dichloropropane-d6	6.33	67	124707	5.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	115.60%
41) Toluene-d8	7.56	98	280478	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
43) trans-1,3-Dichloropropene-	7.85	79	40675	5.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.60%
46) 2-Hexanone-d5	8.31	63	185790	53.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.64%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	93378	6.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.80%#
64) 1,2-Dichlorobenzene-d4	11.86	152	98540	6.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.40%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU040819\
Data File : VU030880.D
Acq On : 08 Apr 2019 01:01
Operator : JC/SP
Sample : K2284-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZB1

Quant Time: Apr 08 05:56:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 04:40:32 2019
Response via : Initial Calibration

