

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040919\
 Data File : VU030903.D
 Acq On : 08 Apr 2019 14:15
 Operator : JC/SP
 Sample : K2259-13ME
 Misc : 5.89g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETJ33ME

Quant Time: Apr 09 02:56:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 09 02:16:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	626041	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	618154	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	286054	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	223452	39.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.96%
7) Chloroethane-d5	1.64	69	186538	43.98	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.96%
11) 1,1-Dichloroethene-d2	2.23	63	373373	36.94	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.88%
21) 2-Butanone-d5	4.19	46	515538	116.40	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.40%
24) Chloroform-d	4.64	84	457499	53.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.18%
26) 1,2-Dichloroethane-d4	5.30	65	323189	55.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.96%
32) Benzene-d6	5.33	84	940366	52.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.22%
36) 1,2-Dichloropropane-d6	6.32	67	331210	53.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.12%
41) Toluene-d8	7.56	98	808729	50.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.12%
43) trans-1,3-Dichloropropene-	7.85	79	148711	52.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	105.46%
47) 2-Hexanone-d5	8.32	63	380209	134.51	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	134.51%#
57) 1,1,2,2-Tetrachloroethane-	10.43	84	516241	61.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	123.76%#
64) 1,2-Dichlorobenzene-d4	11.86	152	325027	57.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.56%

Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	4.21	96	120570	23.765 ug/L	96
34) Trichloroethene	6.19	95	5018	1.011 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040919\
Data File : VU030903.D
Acq On : 08 Apr 2019 14:15
Operator : JC/SP
Sample : K2259-13ME
Misc : 5.89g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETJ33ME

Quant Time: Apr 09 02:56:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 09 02:16:31 2019
Response via : Initial Calibration

