

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040920\
 Data File : VU037622.D
 Acq On : 09 Apr 2020 16:52
 Operator : JC/MD
 Sample : L1885-03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0ET7

Quant Time: Apr 10 05:15:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 10 03:10:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	602671	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	567987	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	277950	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	215612	40.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.86%
7) Chloroethane-d5	1.93	69	184538	43.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.68%
11) 1,1-Dichloroethene-d2	2.59	63	280258	32.64	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.28%
21) 2-Butanone-d5	4.66	46	385684	90.35	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.35%
24) Chloroform-d	5.10	84	344688	43.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.08%
26) 1,2-Dichloroethane-d4	5.74	65	243353	45.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.48%
32) Benzene-d6	5.76	84	760715	45.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.04%
36) 1,2-Dichloropropane-d6	6.72	67	256891	46.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.36%
41) Toluene-d8	7.92	98	694633	44.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.66%
43) trans-1,3-Dichloropropene-	8.20	79	132299	44.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.00%
47) 2-Hexanone-d5	8.66	63	287787	90.37	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.37%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	347791	44.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.52%
64) 1,2-Dichlorobenzene-d4	12.20	152	255013	44.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.18%


Target Compounds					Ovalue
13) Acetone	2.66	43	12502	4.959 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040920\
Data File : VU037622.D
Acq On : 09 Apr 2020 16:52
Operator : JC/MD
Sample : L1885-03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0ET7

Quant Time: Apr 10 05:15:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 10 03:10:48 2020
Response via : Initial Calibration





#13

Acetone

Concen: 4.959 ug/L

RT: 2.66 min Scan# 409

Delta R.T. 0.00 min

Lab File: VU037622.D

Acq: 09 Apr 2020 16:52

Instrument :

MSVOA_U

ClientSampleId :

C0ET7

Tot Ion: 43 Resp: 12502

Ion Ratio Lower Upper

43 100

58 32.8 0.0 63.2

