

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU040920\
 Data File : VU037623.D
 Acq On : 09 Apr 2020 17:15
 Operator : JC/MD
 Sample : L1885-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0EW3

Quant Time: Apr 10 05:18:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040920WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 10 03:10:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	557630	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	528724	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	254461	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	225777	46.32	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.64%
7) Chloroethane-d5	1.93	69	191226	48.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.08%
11) 1,1-Dichloroethene-d2	2.59	63	285418	35.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.86%
21) 2-Butanone-d5	4.66	46	382898	96.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.94%
24) Chloroform-d	5.10	84	354418	48.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.76%
26) 1,2-Dichloroethane-d4	5.74	65	250328	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.60%
32) Benzene-d6	5.76	84	784553	50.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.88%
36) 1,2-Dichloropropane-d6	6.72	67	261963	50.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	101.18%
41) Toluene-d8	7.92	98	709335	49.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.36%
43) trans-1,3-Dichloropropene-	8.20	79	134294	47.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.96%
47) 2-Hexanone-d5	8.65	63	291726	98.41	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.41%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	357878	48.93	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.86%
64) 1,2-Dichlorobenzene-d4	12.20	152	262014	49.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.96%


Target Compounds					Ovalue
13) Acetone	2.66	43	14543	6.235 ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU040920\
Data File : VU037623.D
Acq On : 09 Apr 2020 17:15
Operator : JC/MD
Sample : L1885-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0EW3

Quant Time: Apr 10 05:18:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040920WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 10 03:10:48 2020
Response via : Initial Calibration





#13

Acetone

Concen: 6.235 ug/L

RT: 2.66 min Scan# 409

Delta R.T. 0.00 min

Lab File: VU037623.D

Acq: 09 Apr 2020 17:15

Instrument :

MSVOA_U

ClientSampleId :

C0EW3

Tot Ion: 43 Resp: 14543

Ion Ratio Lower Upper

43 100

58 33.7 0.0 63.2

