

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040921\
 Data File : VU043008.D
 Acq On : 09 Apr 2021 17:27
 Operator : SY/MD
 Sample : M1961-22
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TR2-D-BOTTOM

Quant Time: Apr 10 05:07:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U040921W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 09 09:26:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.382	168	122786	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.256	114	202699	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.423	117	203246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.819	152	98546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.710	65	100718	50.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.36%	
35) Dibromofluoromethane	5.298	113	71616	48.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.76%	
50) Toluene-d8	7.906	98	263305	49.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.76%	
62) 4-Bromofluorobenzene	10.639	95	98006	44.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.98%	
Target Compounds						
						Qvalue
16) Acetone	2.629	43	16600	14.75	ug/l	98
20) Methylene Chloride	3.054	84	6752	4.69	ug/l	94
43) Isopropyl Acetate	5.919	43	13971	2.81	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040921\
Data File : VU043008.D
Acq On : 09 Apr 2021 17:27
Operator : SY/MD
Sample : M1961-22
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TR2-D-BOTTOM

Quant Time: Apr 10 05:07:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U040921W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 09 09:26:47 2021
Response via : Initial Calibration

