



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040921\
Data File : VU042997.D
Acq On : 09 Apr 2021 11:21
Operator : SY/MD
Sample : BFB
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\82U040921W.M
Title : SW846 8260
Last Update : Fri Apr 09 09:26:47 2021



AutoFind: Scans 2892, 2893, 2894; Background Corrected with Scan 2880

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.6	3526	PASS
75	95	30	60	59.5	9282	PASS
95	95	100	100	100.0	15600	PASS
96	95	5	9	6.8	1054	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.2	13604	PASS
175	174	5	9	8.2	1118	PASS
176	174	95	101	99.1	13482	PASS
177	176	5	9	6.3	852	PASS