

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040922\
 Data File : VU047880.D
 Acq On : 08 Apr 2022 20:28
 Operator : SY/MD
 Sample : N2343-10
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0J09

Quant Time: Apr 09 02:07:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 09 01:58:44 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.253	114	287419	50.000	ug/L	0.00
28) Chlorobenzene-d5		9.420	117	305907	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.812	152	156214	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.600	65	107423	51.540	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	103.080%	
7) Chloroethane-d5		1.916	69	93376	56.817	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	113.640%	
11) 1,1-Dichloroethene-d2		2.568	63	140654	36.885	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.760%	
21) 2-Butanone-d5		4.620	46	168930	94.429	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.430%	
24) Chloroform-d		5.067	84	193880	50.838	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.680%	
26) 1,2-Dichloroethane-d4		5.703	65	124169	53.087	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.180%	
32) Benzene-d6		5.729	84	391282	50.832	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.660%	
36) 1,2-Dichloropropane-d6		6.693	67	126487	53.506	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.020%	
41) Toluene-d8		7.899	98	345874	46.119	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.240%	
43) trans-1,3-Dichloroprop...		8.179	79	52349	43.710	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.420%	
47) 2-Hexanone-d5		8.632	63	118657	85.743	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.740%	
56) 1,1,2,2-Tetrachloroeth...		10.758	84	194199	47.868	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.740%	
66) 1,2-Dichlorobenzene-d4		12.195	152	148178	53.317	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.640%	
Target Compounds							
51) Chlorobenzene		9.449	112	4207	0.638	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040922\
Data File : VU047880.D
Acq On : 08 Apr 2022 20:28
Operator : SY/MD
Sample : N2343-10
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0J09

Quant Time: Apr 09 02:07:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 09 01:58:44 2022
Response via : Initial Calibration

