

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040922\
 Data File : VU047872.D
 Acq On : 08 Apr 2022 17:24
 Operator : SY/MD
 Sample : VU0409WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK080

Quant Time: Apr 09 02:05:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 09 01:58:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	306489	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	320467	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	162971	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	115964	52.176	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 104.360%		
7) Chloroethane-d5	1.916	69	99757	56.923	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 113.840%		
11) 1,1-Dichloroethene-d2	2.568	63	149462	36.756	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 73.520%		
21) 2-Butanone-d5	4.620	46	180534	94.636	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 94.640%		
24) Chloroform-d	5.066	84	210458	51.752	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.500%		
26) 1,2-Dichloroethane-d4	5.703	65	134367	53.872	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 107.740%		
32) Benzene-d6	5.729	84	417517	51.776	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 103.560%		
36) 1,2-Dichloropropane-d6	6.693	67	133981	54.101	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 108.200%		
41) Toluene-d8	7.899	98	374385	47.652	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 95.300%		
43) trans-1,3-Dichloroprop...	8.182	79	57495	45.826	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 91.660%		
47) 2-Hexanone-d5	8.632	63	129085	89.041	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 89.040%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	203532	47.889	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 95.780%		
66) 1,2-Dichlorobenzene-d4	12.195	152	153879	53.072	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 106.140%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU040922\
Data File : VU047872.D
Acq On : 08 Apr 2022 17:24
Operator : SY/MD
Sample : VU0409WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK080

Quant Time: Apr 09 02:05:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 09 01:58:44 2022
Response via : Initial Calibration

