

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030919.D
 Acq On : 08 Apr 2019 20:56
 Operator : JC/SP
 Sample : K2284-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZA9

Quant Time: Apr 09 02:11:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 09 01:11:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	253875	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	259157	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	102277	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	102814	5.30	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.00%
7) Chloroethane-d5	1.68	69	89694	5.73	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.60%
11) 1,1-Dichloroethene-d2	2.27	63	154759	3.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.00%
20) 2-Butanone-d5	4.19	46	328377	65.69	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	131.38%#
24) Chloroform-d	4.64	84	160864	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.80%
26) 1,2-Dichloroethane-d4	5.31	65	106984	6.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	122.60%
32) Benzene-d6	5.33	84	356644	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
36) 1,2-Dichloropropane-d6	6.33	67	123708	5.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.00%
41) Toluene-d8	7.56	98	278289	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
43) trans-1,3-Dichloropropene-	7.85	79	42785	5.03	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.60%
46) 2-Hexanone-d5	8.31	63	218372	59.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.16%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	98921	6.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	120.60%#
64) 1,2-Dichlorobenzene-d4	11.86	152	102077	5.97	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.40%

Target Compounds					Ovalue
13) Acetone	2.32	43	11170	2.890 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030919.D
 Acq On : 08 Apr 2019 20:56
 Operator : JC/SP
 Sample : K2284-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZA9

Quant Time: Apr 09 02:11:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 09 01:11:42 2019
 Response via : Initial Calibration

