

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030920.D
 Acq On : 08 Apr 2019 21:21
 Operator : JC/SP
 Sample : K2284-13
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZB1

Quant Time: Apr 09 02:16:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 09 01:11:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	247011	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	251984	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	100950	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	102142	5.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.20%
7) Chloroethane-d5	1.68	69	89367	5.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.40%
11) 1,1-Dichloroethene-d2	2.27	63	154759	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.20%
20) 2-Butanone-d5	4.19	46	321550	66.11	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	132.22%#
24) Chloroform-d	4.64	84	159244	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.40%
26) 1,2-Dichloroethane-d4	5.31	65	107578	6.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	126.60%
32) Benzene-d6	5.33	84	355110	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
36) 1,2-Dichloropropane-d6	6.33	67	123426	5.54	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.80%
41) Toluene-d8	7.56	98	277505	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
43) trans-1,3-Dichloropropene-	7.85	79	41126	4.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.60%
46) 2-Hexanone-d5	8.31	63	215322	59.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	119.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	101591	6.37	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	127.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	100650	5.96	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.20%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU030920.D
Acq On : 08 Apr 2019 21:21
Operator : JC/SP
Sample : K2284-13
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZB1

Quant Time: Apr 09 02:16:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 09 01:11:42 2019
Response via : Initial Calibration

