

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030924.D
 Acq On : 08 Apr 2019 22:58
 Operator : JC/SP
 Sample : K2284-19
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZB7

Quant Time: Apr 09 05:17:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 09 01:11:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	222748	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	226185	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	89909	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	94282	5.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.80%
7) Chloroethane-d5	1.68	69	82996	6.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.00%
11) 1,1-Dichloroethene-d2	2.27	63	142582	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.80%
20) 2-Butanone-d5	4.19	46	306403	69.86	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	139.72%#
24) Chloroform-d	4.64	84	148763	5.15	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.00%
26) 1,2-Dichloroethane-d4	5.31	65	99054	6.47	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	129.40%
32) Benzene-d6	5.33	84	326772	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
36) 1,2-Dichloropropane-d6	6.33	67	112475	5.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.60%
41) Toluene-d8	7.56	98	253933	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	7.85	79	37669	5.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.60%
46) 2-Hexanone-d5	8.31	63	197596	61.25	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.50%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	92590	6.47	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	129.40%#
64) 1,2-Dichlorobenzene-d4	11.86	152	89135	5.93	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.60%

Target Compounds					Ovalue
13) Acetone	2.32	43	12952	3.819 ug/L	96
34) Trichloroethene	6.18	95	138182	7.214 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU030924.D
Acq On : 08 Apr 2019 22:58
Operator : JC/SP
Sample : K2284-19
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZB7

Quant Time: Apr 09 05:17:50 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 09 01:11:42 2019
Response via : Initial Calibration

