

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030962.D
 Acq On : 09 Apr 2019 17:20
 Operator : JC/SP
 Sample : K2283-11
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZ89

Quant Time: Apr 10 03:58:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 01:53:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	233569	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	239516	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	94148	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	73766	4.13	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.60%
7) Chloroethane-d5	1.68	69	68277	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.80%
11) 1,1-Dichloroethene-d2	2.27	63	116424	3.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	63.80%
20) 2-Butanone-d5	4.19	46	238284	51.81	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.62%
24) Chloroform-d	4.64	84	126568	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.60%
26) 1,2-Dichloroethane-d4	5.31	65	79160	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.33	84	260038	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.60%
36) 1,2-Dichloropropane-d6	6.33	67	90540	4.28	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.60%
41) Toluene-d8	7.56	98	207748	3.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.80%
43) trans-1,3-Dichloropropene-	7.85	79	32000	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	8.31	63	143678	42.06	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.12%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	75564	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	71586	4.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU030962.D
Acq On : 09 Apr 2019 17:20
Operator : JC/SP
Sample : K2283-11
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZ89

Quant Time: Apr 10 03:58:59 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 01:53:07 2019
Response via : Initial Calibration

