

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030998.D
 Acq On : 10 Apr 2019 08:45
 Operator : JC/SP
 Sample : VIBLK40
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK40

Quant Time: Apr 11 01:36:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 11 01:24:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	222391	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	230100	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	89728	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	76351	4.49	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	89.80%
7) Chloroethane-d5	1.68	69	72345	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	105.60%
11) 1,1-Dichloroethene-d2	2.27	63	120689	3.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.40%
20) 2-Butanone-d5	4.19	46	237719	54.28	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.56%
24) Chloroform-d	4.65	84	125535	4.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.00%
26) 1,2-Dichloroethane-d4	5.31	65	84105	5.50	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.00%
32) Benzene-d6	5.34	84	274455	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
36) 1,2-Dichloropropane-d6	6.33	67	95356	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.80%
41) Toluene-d8	7.56	98	207812	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
43) trans-1,3-Dichloropropene-	7.85	79	27822	3.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.80%
46) 2-Hexanone-d5	8.31	63	146121	44.52	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	74972	5.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	76891	5.12	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.40%

Target Compounds					Ovalue
16) Methylene chloride	2.70	84	4056	0.214 ug/L	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU030998.D
Acq On : 10 Apr 2019 08:45
Operator : JC/SP
Sample : VIBLK40
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK40

Quant Time: Apr 11 01:36:27 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 11 01:24:42 2019
Response via : Initial Calibration

