

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030917.D
 Acq On : 08 Apr 2019 20:07
 Operator : JC/SP
 Sample : K2284-05
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZA3

Quant Time: Apr 09 01:47:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Apr 09 01:11:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	257931	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	263389	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	105750	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	105819	5.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.40%
7) Chloroethane-d5	1.68	69	90083	5.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.40%
11) 1,1-Dichloroethene-d2	2.27	63	158166	3.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.40%
20) 2-Butanone-d5	4.19	46	332883	65.54	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	131.08%#
24) Chloroform-d	4.64	84	167191	5.00	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.00%
26) 1,2-Dichloroethane-d4	5.31	65	112676	6.35	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	127.00%
32) Benzene-d6	5.33	84	361222	5.16	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.20%
36) 1,2-Dichloropropane-d6	6.33	67	126408	5.43	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.60%
41) Toluene-d8	7.56	98	290892	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
43) trans-1,3-Dichloropropene-	7.85	79	41234	4.77	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.40%
46) 2-Hexanone-d5	8.31	63	230709	61.41	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	122.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	102315	6.14	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	122.80%#
64) 1,2-Dichlorobenzene-d4	11.86	152	100530	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU030917.D
Acq On : 08 Apr 2019 20:07
Operator : JC/SP
Sample : K2284-05
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZA3

Quant Time: Apr 09 01:47:38 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Apr 09 01:11:42 2019
Response via : Initial Calibration

