

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU030957.D
 Acq On : 09 Apr 2019 14:55
 Operator : JC/SP
 Sample : K2286-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BEZF5

Quant Time: Apr 10 03:56:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 01:53:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	218287	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	224506	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	87497	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	81008	4.86	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.20%
7) Chloroethane-d5	1.68	69	76901	5.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	114.40%
11) 1,1-Dichloroethene-d2	2.27	63	129110	3.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.60%
20) 2-Butanone-d5	4.19	46	260252	60.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.10%
24) Chloroform-d	4.64	84	139890	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.31	65	89894	5.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	119.80%
32) Benzene-d6	5.33	84	289421	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
36) 1,2-Dichloropropane-d6	6.33	67	101556	5.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.40%
41) Toluene-d8	7.56	98	223133	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.60%
43) trans-1,3-Dichloropropene-	7.85	79	34027	4.62	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.40%
46) 2-Hexanone-d5	8.31	63	167541	52.32	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.64%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	84913	5.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	119.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	80700	5.51	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU030957.D
Acq On : 09 Apr 2019 14:55
Operator : JC/SP
Sample : K2286-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BEZF5

Quant Time: Apr 10 03:56:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 01:53:07 2019
Response via : Initial Calibration

