

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
 Data File : VU031010.D
 Acq On : 10 Apr 2019 14:43
 Operator : JC/SP
 Sample : K2290-18
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Apr 11 02:52:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 11 02:07:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	229831	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	235329	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	95238	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	73039	4.16	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.20%
7) Chloroethane-d5	1.68	69	71118	5.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.40%
11) 1,1-Dichloroethene-d2	2.27	63	121467	3.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.60%
20) 2-Butanone-d5	4.18	46	257269	56.85	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.70%
24) Chloroform-d	4.64	84	138083	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	5.31	65	89702	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.40%
32) Benzene-d6	5.33	84	281249	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.33	67	96884	4.66	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.20%
41) Toluene-d8	7.56	98	215687	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloropropene-	7.85	79	32895	4.26	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.20%
46) 2-Hexanone-d5	8.31	63	165651	49.35	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.70%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	83233	5.59	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	83792	5.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.20%


Target Compounds					Ovalue
16) Methylene chloride	2.69	84	4291	0.219 ug/L	87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041019\
Data File : VU031010.D
Acq On : 10 Apr 2019 14:43
Operator : JC/SP
Sample : K2290-18
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Apr 11 02:52:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR032919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 11 02:07:02 2019
Response via : Initial Calibration





#16
 Methvlene chloride
 Concen: 0.219 ug/L
 RT: 2.69 min Scan# 499
 Delta R.T. -0.00 min
 Lab File: VU031010.D
 Acq: 10 Apr 2019 14:43

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Tot Ion: 84 Resp: 4291
 Ion Ratio Lower Upper
 84 100
 86 59.0 43.4 80.6
 49 119.9 98.5 182.9

