

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
 Data File : VU058570.D
 Acq On : 10 Apr 2024 19:16
 Operator : MD/SY
 Sample : P2013-21
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0MC5

Quant Time: Apr 11 01:34:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 11 01:28:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	97620	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	96359	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	40217	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	22192	3.434	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.600%
7) Chloroethane-d5	1.907	69	20436	3.638	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.800%
11) 1,1-Dichloroethene-d2	2.560	65	10189	3.483	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.600%
20) 2-Butanone-d5	4.621	46	45891	33.693	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	67.380%
24) Chloroform-d	5.052	84	44524	3.451	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	69.000%#
26) 1,2-Dichloroethane-d4	5.695	65	21818	3.624	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	72.400%
32) Benzene-d6	5.717	84	88063	3.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	67.800%#
36) 1,2-Dichloropropane-d6	6.682	67	28876	3.549	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	71.000%
41) Toluene-d8	7.891	98	73566	3.198	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	64.000%#
43) trans-1,3-Dichloroprop...	8.177	79	7763	3.074	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.400%
46) 2-Hexanone-d5	8.630	63	29654	28.469	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	56.940%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	17160	2.988	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	59.800%#
66) 1,2-Dichlorobenzene-d4	12.190	152	22202	3.278	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	65.600%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
Data File : VU058570.D
Acq On : 10 Apr 2024 19:16
Operator : MD/SY
Sample : P2013-21
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0MC5

Quant Time: Apr 11 01:34:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 11 01:28:06 2024
Response via : Initial Calibration

