

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
 Data File : VU058580.D
 Acq On : 11 Apr 2024 01:26
 Operator : MD/SY
 Sample : P2014-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0MD1

Quant Time: Apr 11 03:00:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 11 02:34:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	79245	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	78633	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	33988	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	22606	4.309	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.904	69	20071	4.402	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	2.560	65	10713	4.511	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.200%
20) 2-Butanone-d5	4.618	46	45822	41.443	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.880%
24) Chloroform-d	5.052	84	46172	4.408	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	5.695	65	22029	4.507	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.200%
32) Benzene-d6	5.721	84	91068	4.302	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.000%
36) 1,2-Dichloropropane-d6	6.682	67	28717	4.325	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.600%
41) Toluene-d8	7.894	98	74521	3.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	8.177	79	7811	3.790	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.800%
46) 2-Hexanone-d5	8.631	63	25344	29.816	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	59.640%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	17559	3.746	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	75.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	22980	4.014	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	80.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
Data File : VU058580.D
Acq On : 11 Apr 2024 01:26
Operator : MD/SY
Sample : P2014-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0MD1

Quant Time: Apr 11 03:00:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 11 02:34:51 2024
Response via : Initial Calibration

