

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
 Data File : VU058589.D
 Acq On : 11 Apr 2024 04:58
 Operator : MD/SY
 Sample : P2035-01 200X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AN7

Quant Time: Apr 11 06:50:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 11 02:34:51 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	88018	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	83714	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	34346	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	20939	3.593	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.800%
7) Chloroethane-d5	1.907	69	20222	3.993	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.800%
11) 1,1-Dichloroethene-d2	2.560	65	9880	3.746	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.000%
20) 2-Butanone-d5	4.624	46	47504	38.682	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.360%
24) Chloroform-d	5.052	84	44994	3.868	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.400%
26) 1,2-Dichloroethane-d4	5.695	65	22327	4.113	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.200%
32) Benzene-d6	5.717	84	86007	3.816	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.400%
36) 1,2-Dichloropropane-d6	6.682	67	28525	4.036	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.800%
41) Toluene-d8	7.891	98	68075	3.406	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.200%#
43) trans-1,3-Dichloroprop...	8.177	79	7970	3.633	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	72.600%
46) 2-Hexanone-d5	8.631	63	32717	36.154	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.300%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	18006	3.609	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.200%
66) 1,2-Dichlorobenzene-d4	12.187	152	20564	3.555	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	71.000%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
Data File : VU058589.D
Acq On : 11 Apr 2024 04:58
Operator : MD/SY
Sample : P2035-01 200X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COAN7

Quant Time: Apr 11 06:50:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 11 02:34:51 2024
Response via : Initial Calibration

