

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
 Data File : VU058557.D
 Acq On : 10 Apr 2024 14:11
 Operator : MD/SY
 Sample : VIBLK199
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VIBLK199

Quant Time: Apr 11 01:31:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 11 01:28:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	61712	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	60120	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	25944	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	22867	5.597	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.000%
7) Chloroethane-d5	1.904	69	19838	5.587	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.800%
11) 1,1-Dichloroethene-d2	2.560	65	10386	5.616	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	112.400%
20) 2-Butanone-d5	4.628	46	43438	50.449	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.900%
24) Chloroform-d	5.052	84	43184	5.295	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.800%
26) 1,2-Dichloroethane-d4	5.695	65	21430	5.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.600%
32) Benzene-d6	5.718	84	86574	5.349	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
36) 1,2-Dichloropropane-d6	6.682	67	27689	5.455	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.000%
41) Toluene-d8	7.891	98	70870	4.937	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.800%
43) trans-1,3-Dichloroprop...	8.177	79	7324	4.648	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.000%
46) 2-Hexanone-d5	8.631	63	28073	43.197	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.400%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	16227	4.528	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.600%
66) 1,2-Dichlorobenzene-d4	12.190	152	20322	4.651	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041024\
Data File : VU058557.D
Acq On : 10 Apr 2024 14:11
Operator : MD/SY
Sample : VIBLK199
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VIBLK199

Quant Time: Apr 11 01:31:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 11 01:28:06 2024
Response via : Initial Calibration

