

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031021.D
 Acq On : 10 Apr 2019 19:12
 Operator : JC/SP
 Sample : K2249-15DL 50X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E3YE7DL

Quant Time: Apr 11 06:39:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 11 05:17:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	451021	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	453957	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	201415	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	220439	54.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	109.50%
7) Chloroethane-d5	1.68	69	175953	57.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	115.16%
11) 1,1-Dichloroethene-d2	2.27	63	311251	42.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.48%
21) 2-Butanone-d5	4.17	46	358035	112.20	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.20%
24) Chloroform-d	4.64	84	343138	55.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.58%
26) 1,2-Dichloroethane-d4	5.30	65	245779	58.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	117.14%
32) Benzene-d6	5.33	84	723233	54.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.14%
36) 1,2-Dichloropropane-d6	6.32	67	245836	54.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.26%
41) Toluene-d8	7.56	98	622179	52.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.94%
43) trans-1,3-Dichloropropene-	7.85	79	107165	51.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	103.50%
47) 2-Hexanone-d5	8.31	63	207339	99.89	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.89%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	338526	55.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	110.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	229709	57.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.98%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.40	62	25571	5.155	ug/L #	72
8) Chloroethane	1.69	64	13052	4.613	ug/L	100
12) 1,1-Dichloroethene	2.28	96	9432	3.025	ug/L #	1
17) trans-1,2-Dichloroethene	2.98	96	6764	2.043	ug/L	93
19) 1,1-Dichloroethane	3.44	63	307095	44.953	ug/L	96
20) cis-1,2-Dichloroethene	4.22	96	453640	124.111	ug/L	97
30) 1,1,1-Trichloroethane	4.91	97	352891	63.787	ug/L	100
34) Trichloroethene	6.19	95	3955	1.085	ug/L	88
42) Toluene	7.64	91	15504	1.022	ug/L	98
52) Ethylbenzene	9.25	91	29373	1.777	ug/L	95
53) m,p-Xylene	9.37	106	25217	4.374	ug/L	96
54) o-xylene	9.78	106	26245	4.577	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031021.D
Acq On : 10 Apr 2019 19:12
Operator : JC/SP
Sample : K2249-15DL 50X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E3YE7DL

Quant Time: Apr 11 06:39:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 11 05:17:40 2019
Response via : Initial Calibration

