

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031068.D
 Acq On : 11 Apr 2019 15:05
 Operator : JC/SP
 Sample : VU0411WBL03
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK56

Quant Time: Apr 12 05:04:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 12 04:43:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	467662	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	458202	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	199783	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	213164	51.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.12%
7) Chloroethane-d5	1.68	69	160715	50.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	101.44%
11) 1,1-Dichloroethene-d2	2.27	63	285698	37.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.68%
21) 2-Butanone-d5	4.17	46	308834	93.34	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.34%
24) Chloroform-d	4.64	84	316395	49.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.22%
26) 1,2-Dichloroethane-d4	5.31	65	223490	51.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.72%
32) Benzene-d6	5.34	84	659150	49.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.56%
36) 1,2-Dichloropropane-d6	6.32	67	221566	48.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.66%
41) Toluene-d8	7.56	98	578248	48.77	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.54%
43) trans-1,3-Dichloropropene-	7.85	79	100859	48.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.50%
47) 2-Hexanone-d5	8.31	63	189941	90.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.66%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	300736	48.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.28%
64) 1,2-Dichlorobenzene-d4	11.86	152	206831	52.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.28%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031068.D
Acq On : 11 Apr 2019 15:05
Operator : JC/SP
Sample : VU0411WBL03
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 59 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK56

Quant Time: Apr 12 05:04:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 12 04:43:47 2019
Response via : Initial Calibration

