

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031014.D
 Acq On : 10 Apr 2019 16:33
 Operator : JC/SP
 Sample : VU0411MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK52

Quant Time: Apr 11 05:29:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 11 05:17:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	448317	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	441590	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	180414	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	192364	48.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	96.14%
7) Chloroethane-d5	1.68	69	155575	51.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	102.44%
11) 1,1-Dichloroethene-d2	2.27	63	269014	37.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.32%
21) 2-Butanone-d5	4.17	46	293528	92.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.54%
24) Chloroform-d	4.64	84	324489	53.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.16%
26) 1,2-Dichloroethane-d4	5.30	65	221364	53.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.14%
32) Benzene-d6	5.33	84	631277	48.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.94%
36) 1,2-Dichloropropane-d6	6.32	67	220312	49.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.74%
41) Toluene-d8	7.56	98	536175	46.92	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.84%
43) trans-1,3-Dichloropropene-	7.85	79	94518	46.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.84%
47) 2-Hexanone-d5	8.31	63	177260	87.79	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.79%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	309051	51.86	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.72%
64) 1,2-Dichlorobenzene-d4	11.86	152	193484	54.53	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.06%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031014.D
Acq On : 10 Apr 2019 16:33
Operator : JC/SP
Sample : VU0411MBL01
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK52

Quant Time: Apr 11 05:29:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 11 05:17:40 2019
Response via : Initial Calibration

