

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031058.D
 Acq On : 11 Apr 2019 10:43
 Operator : JC/SP
 Sample : K2341-05
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB716

Quant Time: Apr 12 04:13:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 12 03:03:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	472004	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	470433	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	205356	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	205063	48.67	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.34%
7) Chloroethane-d5	1.68	69	173082	54.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.24%
11) 1,1-Dichloroethene-d2	2.27	63	281983	37.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.00%
21) 2-Butanone-d5	4.17	46	355715	106.52	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.52%
24) Chloroform-d	4.64	84	357632	55.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.12%
26) 1,2-Dichloroethane-d4	5.31	65	245857	55.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.96%
32) Benzene-d6	5.34	84	707124	51.49	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.98%
36) 1,2-Dichloropropane-d6	6.32	67	240079	51.01	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.02%
41) Toluene-d8	7.56	98	617356	50.72	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.44%
43) trans-1,3-Dichloropropene-	7.85	79	99518	46.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.74%
47) 2-Hexanone-d5	8.31	63	220134	102.34	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.34%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	329034	51.83	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.66%
64) 1,2-Dichlorobenzene-d4	11.86	152	227601	56.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.72%

Target Compounds


					Ovalue
13) Acetone	2.32	43	20253	5.927 ug/L	99
25) Chloroform	4.67	83	80485	11.201 ug/L	99
38) Bromodichloromethane	6.76	83	30569	5.813 ug/L	99
42) Toluene	7.64	91	9027	0.574 ug/L	98
46) Tetrachloroethene	8.22	164	1427459	520.000 ug/L	99
49) Dibromochloromethane	8.48	129	13208	3.367 ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031058.D
Acq On : 11 Apr 2019 10:43
Operator : JC/SP
Sample : K2341-05
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB716

Quant Time: Apr 12 04:13:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 12 03:03:54 2019
Response via : Initial Calibration





#13

Acetone

Concen: 5.927 ug/L

RT: 2.32 min Scan# 382

Delta R.T. 0.00 min

Lab File: VU031058.D

Acq: 11 Apr 2019 10:43

Instrument :

MSVOA_U

ClientSampleId :


DB716

Tot Ion: 43 Resp: 20253

Ion Ratio Lower Upper

43 100

58 31.3 0.0 64.0



Abundance Ion 43.00 (42.70 to 43.70): VU031039.D (-362) (-)

Ion 58.00 (57.70 to 58.70): VU031039.D (-362) (-)

12000

10000

8000

6000



4000

2000

0

2.25 2.30 2.35

Time-->



#25

Chloroform

Concen: 11.201 ug/L

RT: 4.67 min Scan# 1113

Delta R.T. 0.00 min

Lab File: VU031058.D


Acq: 11 Apr 2019 10:43

Tgt Ion: 83 Resp: 80485

Ion Ratio Lower Upper

83 100

85 66.4 45.8 85.0



Abundance Ion 83.00 (82.70 to 83.70): VU031039.D (-1096) (-)

Ion 85.00 (84.70 to 85.70): VU031039.D (-1096) (-)

30000


20000


10000

0

4.60 4.65 4.70 4.75

Time-->





#38

Bromodichloromethane

Concen: 5.813 ug/L

RT: 6.76 min Scan# 1762

Delta R.T. 0.00 min

Lab File: VU031058.D

Acq: 11 Apr 2019 10:43

Instrument :

MSVOA_U

ClientSampleId :

DB716


Tot Ion: 83 Resp: 30569

Ion Ratio Lower Upper

83 100

85 66.6 46.5 86.3




127 9.7 6.6 10.0



Abundance Ion 83.00 (82.70 to 83.70): VU031039.D (-1743) (-)

Ion 85.00 (84.70 to 85.70): VU031039.D (-1743) (-)

Ion 127.00 (126.70 to 127.70): VU031039.D (-1743) (-)



#42

Toluene

Concen: 0.574 ug/L

RT: 7.64 min Scan# 2037

Delta R.T. 0.01 min

Lab File: VU031058.D


Acq: 11 Apr 2019 10:43

Tgt Ion: 91 Resp: 9027

Ion Ratio Lower Upper



91 100


92 56.0 40.0 74.4



Abundance Ion 91.00 (90.70 to 91.70): VU031039.D (-2020) (-)

Ion 92.00 (91.70 to 92.70): VU031039.D (-2020) (-)





#46

Tetrachloroethene

Concen: 520.000 ug/L

RT: 8.22 min Scan# 2219

Delta R.T. -0.00 min

Lab File: VU031058.D

Acq: 11 Apr 2019 10:43

Instrument :

MSVOA_U

ClientSampleId :

DB716

Tot Ion:164 Resp: 1427459


Ion Ratio Lower Upper

164 100

129 96.2 67.4 125.2

131 93.5 65.9 122.5

166 125.7 88.6 164.6






Abundance Ion 164.00 (163.70 to 164.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 131.00 (130.70 to 131.70): V

Ion 166.00 (165.70 to 166.70): V



#49

Dibromochloromethane

Concen: 3.367 ug/L

RT: 8.48 min Scan# 2297

Delta R.T. 0.00 min

Lab File: VU031058.D


Acq: 11 Apr 2019 10:43

Tgt Ion:129 Resp: 13208

Ion Ratio Lower Upper

129 100

127 78.5 54.5 101.3



Abundance Ion 129.00 (128.70 to 129.70): V

Ion 127.00 (126.70 to 127.70): V

