

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031067.D
 Acq On : 11 Apr 2019 14:43
 Operator : JC/SP
 Sample : VU0411MBL02
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK55

Quant Time: Apr 12 05:01:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 12 04:43:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	505832	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	483756	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	213851	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	230173	50.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.94%
7) Chloroethane-d5	1.68	69	178364	52.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.10%
11) 1,1-Dichloroethene-d2	2.27	63	298581	36.56	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.12%
21) 2-Butanone-d5	4.17	46	306362	85.61	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.61%
24) Chloroform-d	4.64	84	346513	50.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.48%
26) 1,2-Dichloroethane-d4	5.30	65	231025	49.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.18%
32) Benzene-d6	5.33	84	701521	49.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.34%
36) 1,2-Dichloropropane-d6	6.32	67	231084	47.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.50%
41) Toluene-d8	7.56	98	610866	48.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.60%
43) trans-1,3-Dichloropropene-	7.85	79	105360	47.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.48%
47) 2-Hexanone-d5	8.31	63	187646	84.83	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.83%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	306491	46.95	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.90%
64) 1,2-Dichlorobenzene-d4	11.86	152	217897	51.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.62%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031067.D
Acq On : 11 Apr 2019 14:43
Operator : JC/SP
Sample : VU0411MBL02
Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 58 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK55

Quant Time: Apr 12 05:01:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 12 04:43:47 2019
Response via : Initial Calibration

