

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031082.D
 Acq On : 11 Apr 2019 20:25
 Operator : JC/SP
 Sample : K2341-07DL 20X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DB722DL

Quant Time: Apr 12 06:53:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 12 04:43:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	458964	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	456001	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	196195	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	157075	38.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.68%
7) Chloroethane-d5	1.68	69	134793	43.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.70%
11) 1,1-Dichloroethene-d2	2.27	63	224000	30.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.46%
21) 2-Butanone-d5	4.17	46	340835	104.96	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.96%
24) Chloroform-d	4.64	84	310790	49.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.32%
26) 1,2-Dichloroethane-d4	5.30	65	216468	50.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.38%
32) Benzene-d6	5.33	84	603384	45.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.66%
36) 1,2-Dichloropropane-d6	6.32	67	214139	46.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.88%
41) Toluene-d8	7.56	98	508726	43.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.22%
43) trans-1,3-Dichloropropene-	7.85	79	93996	45.18	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.36%
47) 2-Hexanone-d5	8.31	63	189140	90.71	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.71%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	309050	50.22	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.44%
64) 1,2-Dichlorobenzene-d4	11.86	152	201217	52.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.30%


Target Compounds					Ovalue
20) cis-1,2-Dichloroethene	4.22	96	76347	20.526 ug/L	93
34) Trichloroethene	6.18	95	41501	11.339 ug/L	93
42) Toluene	7.64	91	19382	1.272 ug/L	96
46) Tetrachloroethene	8.22	164	161867	60.832 ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031082.D
Acq On : 11 Apr 2019 20:25
Operator : JC/SP
Sample : K2341-07DL 20X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 73 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DB722DL

Quant Time: Apr 12 06:53:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 12 04:43:47 2019
Response via : Initial Calibration





#20

cis-1,2-Dichloroethene

Concen: 20.526 ug/L

RT: 4.22 min Scan# 973

Delta R.T. 0.00 min

Lab File: VU031082.D

Acq: 11 Apr 2019 20:25

Instrument :

MSVOA_U

ClientSampleId :

DB722DL


Tot Ion: 96 Resp: 76347

Ion Ratio Lower Upper




96 100

61 138.5 102.9 191.1

68 0.0 0.0 0.0



Abundance Ion 96.00 (95.70 to 96.70): VU031082.D (-880) (-)



#34

Trichloroethene

Concen: 11.339 ug/L

RT: 6.18 min Scan# 1584

Delta R.T. 0.00 min

Lab File: VU031082.D

Acq: 11 Apr 2019 20:25

Tgt Ion: 95 Resp: 41501


Ion Ratio Lower Upper

95 100



97 59.3 44.5 82.7


132 99.0 61.5 114.1

130 97.0 64.9 120.5



Abundance Ion 95.00 (94.70 to 95.70): VU031082.D (-1490) (-)





#42

Toluene

Concen: 1.272 ug/L

RT: 7.64 min Scan# 2037

Delta R.T. 0.01 min

Lab File: VU031082.D

Acq: 11 Apr 2019 20:25

Instrument :

MSVOA_U

ClientSampleId :





DB722DL

Tot Ion: 91 Resp: 19382

Ion Ratio Lower Upper

91 100

92 60.4 40.0 74.4



#46

Tetrachloroethene

Concen: 60.832 ug/L

RT: 8.22 min Scan# 2218

Delta R.T. 0.00 min

Lab File: VU031082.D

Acq: 11 Apr 2019 20:25

Tgt Ion: 164 Resp: 161867

Ion Ratio Lower Upper

164 100

129 97.6 67.4 125.2

131 97.2 65.9 122.5

166 127.4 88.6 164.6

