

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
 Data File : VU031103.D
 Acq On : 12 Apr 2019 04:49
 Operator : JC/SP
 Sample : K2353-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 95 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETJ60

Quant Time: Apr 12 08:21:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 12 07:24:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	497424	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	507041	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	212318	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	163790	36.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	73.78%
7) Chloroethane-d5	1.68	69	137213	40.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.42%
11) 1,1-Dichloroethene-d2	2.27	63	232862	28.99	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.98%#
21) 2-Butanone-d5	4.17	46	339496	96.47	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.47%
24) Chloroform-d	4.64	84	306779	45.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.46%
26) 1,2-Dichloroethane-d4	5.30	65	217494	46.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.98%
32) Benzene-d6	5.33	84	605190	40.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.78%
36) 1,2-Dichloropropane-d6	6.32	67	215660	42.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.02%
41) Toluene-d8	7.56	98	522308	39.81	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.62%#
43) trans-1,3-Dichloropropene-	7.85	79	90761	39.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.48%
47) 2-Hexanone-d5	8.31	63	192384	82.98	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.98%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	315328	46.08	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.16%
64) 1,2-Dichlorobenzene-d4	11.86	152	199237	47.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.42%


Target Compounds					Ovalue
13) Acetone	2.32	43	9909	2.752 ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU041119\
Data File : VU031103.D
Acq On : 12 Apr 2019 04:49
Operator : JC/SP
Sample : K2353-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 95 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETJ60

Quant Time: Apr 12 08:21:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM040519WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 12 07:24:02 2019
Response via : Initial Calibration





#13
 Acetone
 Concen: 2.752 ug/L
 RT: 2.32 min Scan# 381
 Delta R.T. 0.00 min
 Lab File: VU031103.D
 Acq: 12 Apr 2019 04:49

Instrument :
 MSVOA_U
 ClientSampleId :
 ETJ60

Tot Ion: 43 Resp: 9909
 Ion Ratio Lower Upper
 43 100
 58 28.7 0.0 64.0

