

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
 Data File : VU047897.D
 Acq On : 11 Apr 2022 11:52
 Operator : SY/MD
 Sample : VU0411WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK082

Quant Time: Apr 12 02:01:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 12 01:57:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	283712	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	298958	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	157003	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	103659	50.384	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	100.760%
7) Chloroethane-d5	1.916	69	90222	55.616	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	111.240%
11) 1,1-Dichloroethene-d2	2.572	63	136538	36.273	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.540%
21) 2-Butanone-d5	4.623	46	179311	101.541	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.540%
24) Chloroform-d	5.067	84	197764	52.534	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.060%
26) 1,2-Dichloroethane-d4	5.703	65	129130	55.929	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.860%
32) Benzene-d6	5.729	84	397333	52.818	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.640%
36) 1,2-Dichloropropane-d6	6.694	67	127739	55.292	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.580%
41) Toluene-d8	7.899	98	356406	48.628	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.260%
43) trans-1,3-Dichloroprop...	8.182	79	56263	48.071	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.140%
47) 2-Hexanone-d5	8.636	63	130312	96.354	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.350%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	201533	50.830	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.660%
66) 1,2-Dichlorobenzene-d4	12.195	152	154427	55.286	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
Data File : VU047897.D
Acq On : 11 Apr 2022 11:52
Operator : SY/MD
Sample : VU0411WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK082

Quant Time: Apr 12 02:01:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 12 01:57:58 2022
Response via : Initial Calibration

