

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
 Data File : VU047906.D
 Acq On : 11 Apr 2022 15:19
 Operator : SY/MD
 Sample : N2343-05DL 4X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0HY3DL

Quant Time: Apr 12 02:03:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 12 01:57:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	294407	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	310658	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	160947	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	101977	47.766	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.540%
7) Chloroethane-d5	1.916	69	89530	53.184	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	106.360%
11) 1,1-Dichloroethene-d2	2.571	63	133451	34.165	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.340%
21) 2-Butanone-d5	4.620	46	177503	96.866	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.870%
24) Chloroform-d	5.067	84	198544	50.826	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.660%
26) 1,2-Dichloroethane-d4	5.703	65	127706	53.303	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.600%
32) Benzene-d6	5.729	84	395728	50.624	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.240%
36) 1,2-Dichloropropane-d6	6.693	67	126712	52.782	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.560%
41) Toluene-d8	7.899	98	348201	45.719	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.440%
43) trans-1,3-Dichloroprop...	8.179	79	54779	45.040	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.080%
47) 2-Hexanone-d5	8.632	63	122530	87.188	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.190%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	199543	48.433	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.860%
66) 1,2-Dichlorobenzene-d4	12.195	152	151380	52.867	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.740%
Target Compounds						
					Qvalue	
51) Chlorobenzene	9.449	112	327871	48.942	ug/L	98
65) 1,4-Dichlorobenzene	11.841	146	8062	1.610	ug/L	97
67) 1,2-Dichlorobenzene	12.211	146	7760	1.604	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
Data File : VU047906.D
Acq On : 11 Apr 2022 15:19
Operator : SY/MD
Sample : N2343-05DL 4X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0HY3DL

Quant Time: Apr 12 02:03:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 12 01:57:58 2022
Response via : Initial Calibration

