

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
 Data File : VU047924.D
 Acq On : 11 Apr 2022 22:37
 Operator : SY/MD
 Sample : N2293-06ME
 Misc : 6.60g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETNN6ME

Quant Time: Apr 12 05:05:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 12 04:48:51 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		6.250	114	339359	50.000	ug/L	0.00
28) Chlorobenzene-d5		9.423	117	357594	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.815	152	188069	50.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.597	65	197193	80.130	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	= 160.260%#		
7) Chloroethane-d5		1.896	69	127953	65.941	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	= 131.880%#		
11) 1,1-Dichloroethene-d2		2.546	63	276328	61.373	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	= 122.740%		
21) 2-Butanone-d5		4.642	46	428126	202.687	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	= 202.690%#		
24) Chloroform-d		5.063	84	422194	93.762	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 187.520%#		
26) 1,2-Dichloroethane-d4		5.703	65	279569	101.232	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 202.460%#		
32) Benzene-d6		5.722	84	859790	95.553	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	= 191.100%#		
36) 1,2-Dichloropropane-d6		6.690	67	281201	101.760	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	= 203.520%#		
41) Toluene-d8		7.899	98	781691	89.165	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 178.340%#		
43) trans-1,3-Dichloroprop...		8.182	79	119040	85.029	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	= 170.060%#		
47) 2-Hexanone-d5		8.658	63	330506	204.307	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	= 204.310%#		
56) 1,1,2,2-Tetrachloroeth...		10.764	84	453071	95.535	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	= 191.080%#		
66) 1,2-Dichlorobenzene-d4		12.198	152	343203	102.573	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	= 205.140%#		
Target Compounds							Qvalue
33) Benzene		5.771	78	57009	5.508	ug/L	100
42) Toluene		7.970	91	24788	2.183	ug/L	94
51) Chlorobenzene		9.452	112	257015	33.329	ug/L	98
67) 1,2-Dichlorobenzene		12.214	146	4836	0.856	ug/L #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
Data File : VU047924.D
Acq On : 11 Apr 2022 22:37
Operator : SY/MD
Sample : N2293-06ME
Misc : 6.60g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETNN6ME

Quant Time: Apr 12 05:05:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 12 04:48:51 2022
Response via : Initial Calibration

