

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
 Data File : VU047917.D
 Acq On : 11 Apr 2022 19:32
 Operator : SY/MD
 Sample : N2262-14ME
 Misc : 5.31g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW6L4ME

Quant Time: Apr 12 02:05:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 12 01:57:58 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 04/14/2022
 Supervised By : Mahesh Dadoda 04/14/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	529091	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	560850	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.819	152	299582	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	202099	52.674	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 105.340%		
7) Chloroethane-d5	1.896	69	126003m	41.650	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery	= 83.300%		
11) 1,1-Dichloroethene-d2	2.539	63	287171	40.909	ug/L	-0.04
Spiked Amount	50.000	Range 60 - 125	Recovery	= 81.820%		
21) 2-Butanone-d5	4.642	46	414122	125.751	ug/L	0.01
Spiked Amount	100.000	Range 40 - 130	Recovery	= 125.750%		
24) Chloroform-d	5.063	84	420959	59.963	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 119.920%		
26) 1,2-Dichloroethane-d4	5.703	65	278284	64.632	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 129.260%#		
32) Benzene-d6	5.722	84	872245	61.806	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 123.620%		
36) 1,2-Dichloropropane-d6	6.690	67	284070	65.543	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 131.080%#		
41) Toluene-d8	7.899	98	799894	58.175	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 116.340%		
43) trans-1,3-Dichloroprop...	8.185	79	122739	55.899	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 111.800%		
47) 2-Hexanone-d5	8.668	63	330897	130.419	ug/L	0.03
Spiked Amount	100.000	Range 45 - 130	Recovery	= 130.420%#		
56) 1,1,2,2-Tetrachloroeth...	10.767	84	465581	62.594	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 125.180%#		
66) 1,2-Dichlorobenzene-d4	12.198	152	351651	65.978	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 131.960%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041122\
Data File : VU047917.D
Acq On : 11 Apr 2022 19:32
Operator : SY/MD
Sample : N2262-14ME
Misc : 5.31g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW6L4ME

Quant Time: Apr 12 02:05:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM032422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 12 01:57:58 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/14/2022
Supervised By : Mahesh Dadoda 04/14/2022

