

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041123\
 Data File : VU053628.D
 Acq On : 11 Apr 2023 12:18
 Operator : JC/MD
 Sample : VU0411WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK076

Quant Time: Apr 12 02:23:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 12 02:22:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.243	114	230554	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	214922	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	107479	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	68572	38.197	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	76.400%	
7) Chloroethane-d5	1.909	69	61335	44.312	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	88.620%	
11) 1,1-Dichloroethene-d2	2.565	63	85363	30.718	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.440%	
21) 2-Butanone-d5	4.610	46	90052	110.792	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	110.790%	
24) Chloroform-d	5.057	84	128946	43.431	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.860%	
26) 1,2-Dichloroethane-d4	5.697	65	84064	47.916	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	95.840%	
32) Benzene-d6	5.723	84	261658	45.103	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.200%	
36) 1,2-Dichloropropane-d6	6.684	67	83203	46.922	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	93.840%	
41) Toluene-d8	7.893	98	240726	41.749	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	83.500%	
43) trans-1,3-Dichloroprop...	8.176	79	39031	42.642	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	85.280%	
47) 2-Hexanone-d5	8.626	63	67513	111.560	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	111.560%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	123825	49.545	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	99.100%	
66) 1,2-Dichlorobenzene-d4	12.188	152	96645	43.078	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.160%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041123\
Data File : VU053628.D
Acq On : 11 Apr 2023 12:18
Operator : JC/MD
Sample : VU0411WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK076

Quant Time: Apr 12 02:23:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM033123WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 12 02:22:37 2023
Response via : Initial Calibration

