

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU041123\  
 Data File : VU053628.D  
 Acq On : 11 Apr 2023 12:18  
 Operator : JC/MD  
 Sample : VU0411WBL01  
 Misc : 5.0mL/MSVOA\_U/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK076

Quant Time: Apr 12 02:23:12 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM033123WMA.M  
 Quant Title : VOC Analysis  
 QLast Update : Wed Apr 12 02:22:37 2023  
 Response via : Initial Calibration

| Compound                      | R.T.    | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|---------|-------|----------|----------|-------|----------|
| -----                         |         |       |          |          |       |          |
| Internal Standards            |         |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 6.243   | 114   | 230554   | 50.000   | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 9.414   | 117   | 214922   | 50.000   | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.809  | 152   | 107479   | 50.000   | ug/L  | 0.00     |
| System Monitoring Compounds   |         |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.597   | 65    | 68572    | 38.197   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 135 | Recovery | =     | 76.400%  |
| 7) Chloroethane-d5            | 1.909   | 69    | 61335    | 44.312   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 130 | Recovery | =     | 88.620%  |
| 11) 1,1-Dichloroethene-d2     | 2.565   | 63    | 85363    | 30.718   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 61.440%  |
| 21) 2-Butanone-d5             | 4.610   | 46    | 90052    | 110.792  | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 40 - 130 | Recovery | =     | 110.790% |
| 24) Chloroform-d              | 5.057   | 84    | 128946   | 43.431   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 86.860%  |
| 26) 1,2-Dichloroethane-d4     | 5.697   | 65    | 84064    | 47.916   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 95.840%  |
| 32) Benzene-d6                | 5.723   | 84    | 261658   | 45.103   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 125 | Recovery | =     | 90.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.684   | 67    | 83203    | 46.922   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 70 - 120 | Recovery | =     | 93.840%  |
| 41) Toluene-d8                | 7.893   | 98    | 240726   | 41.749   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 83.500%  |
| 43) trans-1,3-Dichloroprop... | 8.176   | 79    | 39031    | 42.642   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 60 - 125 | Recovery | =     | 85.280%  |
| 47) 2-Hexanone-d5             | 8.626   | 63    | 67513    | 111.560  | ug/L  | 0.00     |
| Spiked Amount                 | 100.000 | Range | 45 - 130 | Recovery | =     | 111.560% |
| 56) 1,1,2,2-Tetrachloroeth... | 10.751  | 84    | 123825   | 49.545   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 65 - 120 | Recovery | =     | 99.100%  |
| 66) 1,2-Dichlorobenzene-d4    | 12.188  | 152   | 96645    | 43.078   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000  | Range | 80 - 120 | Recovery | =     | 86.160%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU041123\  
Data File : VU053628.D  
Acq On : 11 Apr 2023 12:18  
Operator : JC/MD  
Sample : VU0411WBL01  
Misc : 5.0mL/MSVOA\_U/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VBLK076

Quant Time: Apr 12 02:23:12 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMULM033123WMA.M  
Quant Title : VOC Analysis  
QLast Update : Wed Apr 12 02:22:37 2023  
Response via : Initial Calibration

