

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041124\
 Data File : VU058634.D
 Acq On : 12 Apr 2024 01:51
 Operator : MD/SY
 Sample : P2055-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COBN5

Quant Time: Apr 12 05:01:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 04:41:05 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	74466	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	73657	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	31043	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	20086	4.074	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.400%
7) Chloroethane-d5	1.908	69	19253	4.494	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.560	65	9915	4.443	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	4.621	46	49333	47.483	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.960%
24) Chloroform-d	5.052	84	46464	4.721	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.400%
26) 1,2-Dichloroethane-d4	5.692	65	22926	4.992	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.800%
32) Benzene-d6	5.718	84	88983	4.487	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	6.682	67	29988	4.822	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.400%
41) Toluene-d8	7.891	98	73661	4.189	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.800%
43) trans-1,3-Dichloroprop...	8.177	79	8183	4.239	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.634	63	34056	42.772	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.540%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	18677	4.254	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	23578	4.510	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	90.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU041124\
Data File : VU058634.D
Acq On : 12 Apr 2024 01:51
Operator : MD/SY
Sample : P2055-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BN5

Quant Time: Apr 12 05:01:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR040424WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 04:41:05 2024
Response via : Initial Calibration

